

**Lakeland Industries**

202 Pride Lane SW
Decatur, Alabama 35603
Telephone: 256-350-3873
Toll Free: 800-645-9291
Fax: 256-350-0773

Certificate of Sterility**Gamma Irradiation for CleanMax Clean Sterile Cleanroom Apparel**

Sterility Lot Number:	CR20180201
Product Release Date:	9/26/2018 00:00
Minimum Specified Dose (kGy):	35.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	36.7
Maximum Delivered Dose (kGy):	46.8

Manufacturing Lot Number:	M18021488
Sterilization Lot Number:	CR20180201

Item numbers:	Size:
CTL428CS	LG, XL, 2X
CTL903CS	LG/XL

Lakeland certifies the product number and lot numbers above have been processed and released according to specifications and standard operating procedures. The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.

Approved by: Lakeland Industries
Quality Assurance