Lakeland Industries
202 Pride Lane SW
Decatur, Alabama 35603
Telephone: 256-350-3873
Toll Free: 800-645-9291
Fax: 256-350-0773

Certificate of Sterility

Gamma Irradiation for CleanMax Clean Sterile Cleanroom Apparel

Sterility Lot Number:	CS20201005US
Product Release Date:	1/20/2021 00:00
Minimum Specified Dose (kGy):	30.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	34.5
Maximum Delivered Dose (kGy):	43.1

Manufacturing Lot Number:	M2100174B
Sterilization Lot Number:	CS20201005US

Item numbers:	Size:
CTL417CS	XL,2X,3X,4X
CTL428CS	LG,XL,2X
CTL903CS	LGXL
CTL850CSP	18"

Lakeland certifies the product number and lot numbers above have been processed and released according to specifications and standard operating procedures. The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Heath Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^-6. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.

Approved by: Lakeland Industries
Quality Assurance

Lakeland Industries
202 Pride Lane SW
Decatur, Alabama 35603
Telephone: 256-350-3873
Toll Free: 800-645-9291
Fax: 256-350-0773

Certificate of Sterility

Gamma Irradiation for CleanMax Clean Sterile Cleanroom Apparel

Sterility Lot Number:	CS20201005US
Product Release Date:	1/13/2021 00:00
Minimum Specified Dose (kGy):	30.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	32.5
Maximum Delivered Dose (kGy):	46.1

Manufacturing Lot Number:	M2003638B
Sterilization Lot Number:	CS20201005US

Item numbers:	Size:
CTL191CS	MD,LG,XL,2X
CTL417CS	MD,LG

Lakeland certifies the product number and lot numbers above have been processed and released according to specifications and standard operating procedures. The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Heath Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^-6. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.

Approved by: Lakeland Industries
Quality Assurance