

**Lakeland Industries**

202 Pride Lane SW
Decatur, Alabama 35603
Telephone: 256-350-3873
Toll Free: 800-645-9291
Fax: 256-350-0773

Certificate of Sterility

Gamma Irradiation for CleanMax Clean Sterile Cleanroom Apparel

Sterility Lot Number:	CS20201228-US
Product Release Date:	3/10/2021 00:00
Minimum Specified Dose (kGy):	30.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	30.0
Maximum Delivered Dose (kGy):	42.3

Manufacturing Lot Number:	M2100211/0212/0213/0214/0215/0216/0200
Sterilization Lot Number:	CS20201228-US

Item numbers:	Size:
CTL903CS	SMMD; 2X
CTL713CS	Unisize
CTL850CSP	18"
CTL191CS	3X; 4X; 5X

Lakeland certifies the product number and lot numbers above have been processed and released according to specifications and standard operating procedures. The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.

Approved by: Lakeland Industries
Quality Assurance

**Lakeland Industries**

202 Pride Lane SW
Decatur, Alabama 35603
Telephone: 256-350-3873
Toll Free: 800-645-9291
Fax: 256-350-0773

Certificate of Sterility**Gamma Irradiation for CleanMax Clean Sterile Cleanroom Apparel**

Sterility Lot Number:	CS20201228-US
Product Release Date:	3/10/2021 00:00
Minimum Specified Dose (kGy):	30.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	30.0
Maximum Delivered Dose (kGy):	41.4

Manufacturing Lot Number:	M2100201/0202/0203/0204/0205/0206/0207/0208/0209/0210
Sterilization Lot Number:	CS20201228-US

Item numbers:	Size:
CTL191CS	MD; LG; XL; 2X
CTL417CS	LG; XL; 3X; 4X; 5X

Lakeland certifies the product number and lot numbers above have been processed and released according to specifications and standard operating procedures. The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.

Approved by: Lakeland Industries
Quality Assurance