



Lakeland Industries

202 Pride Lane SW
 Decatur, Alabama 35603
 Telephone: 256-350-3873
 Toll Free: 800-645-9291
 Fax: 256-350-0773

Certificate of Sterility

Gamma Irradiation for CleanMax Clean Sterile Cleanroom Apparel

Sterility Lot Number:	CS20210106US
Product Release Date:	3/10/2021 00:00
Minimum Specified Dose (kGy):	30.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	30.7
Maximum Delivered Dose (kGy):	45.2

Manufacturing Lot Number:	M2100190/0217/0218/0219/0220	M2100221/0222/0223/0224
Sterilization Lot Number:	CS20210106US	

Item numbers:	Size:
CTL428CS	MD; LG; XL; 2X
CTL903CS	SMMD; LGXL
CTL713CS	Unisize
CTL850CSP	18"

Lakeland certifies the product number and lot numbers above have been processed and released according to specifications and standard operating procedures. The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10⁻⁶. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.

Approved by: Lakeland Industries
 Quality Assurance