

CERTIFICATE OF STERILITY

Gamma Irradiation for CleanMax[®] Clean Sterile Cleanroom Apparel

Sterility Lot Number:	CS20211201-US
Product Release Date:	18MAR2022
Minimum Specified Dose (kGy):	30.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	30.1
Maximum Delivered Dose (kGy):	39.9

Manufacturing Lot Number:	M2200311/0312/0313/0314/0315/0320/0321 M2200304/305/306/307/308/309/310 M2200315/0316/0317/0318/0319
Sterilization Lot Number:	CR20220318

ITEM NUMBER(S)	SIZE(S)
CTL414CS	MD, LG, XL, 3X, 4X, 5X
CTL191CS	MD, LG, XL, 2X, 3X, 4X, 5X
CTL428CS	MD, 2X, 3X, 4X, 5X

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in **ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization** to provide a Sterility Assurance Level of 10⁻⁶. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland Customer Service at info@lakeland.com.



M. LEE MORGAN, Ph.D., Global Director
Quality and Compliance