

## CERTIFICATE OF STERILITY

*Gamma Irradiation for CleanMax<sup>®</sup> Clean Sterile  
Cleanroom Apparel*

<b>Sterility Lot Number:</b>	<b>CS20211202-US</b>
<b>Product Release Date:</b>	20220306
<b>Minimum Specified Dose (kGy):</b>	30.0
<b>Maximum Specified Dose (kGy):</b>	50.0
<b>Minimum Delivered Dose (kGy):</b>	30.3
<b>Maximum Delivered Dose (kGy):</b>	39.9

<b>Manufacturing Lot Number:</b>	M2200238/0239/0240/0241/0242/0243/0244/ 0245
<b>Sterilization Lot Number:</b>	CR20220302

ITEM NUMBER(S)	SIZE(S)
CTL428CS	LG, XL, 5X
CTL713CS	
CTL850CSP	18
CTL903CSP	SMMD, LGXL, 2X

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at [www.lakeland.com](http://www.lakeland.com) or contact Lakeland customer service at [info@lakeland.com](mailto:info@lakeland.com).



**M. LEE MORGAN, Ph.D.**, Global Director  
Quality and Compliance