Lakeland®

Protect Your People®

CERTIFICATE OF STERILITY

Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

Sterility Lot Number:	**CS20220302-US**
Product Release Date:	20220830
Minimum Specified Dose (kGy):	30.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	30.0
Maximum Delivered Dose (kGy):	46.8

Manufacturing Lot Number:	M2201006/1007/1008/1009/1010/1011/1012/1013
Sterilization Lot Number:	CR20220830

ITEM NUMBER(S)	SIZE(S)
CTL903CSP	SMMD, 2X
CTL850CSP	18
CTL191CS	MD,LG,XL,2X

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in **ANSI/AAMI/ISO** 11137-1: Sterilization of Heath Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^-6. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.

M. LEE MORGAN, Ph.D., Global Director
Quality and Compliance

Lakeland Industries, Inc
256-350-3873
800-645-9291
Fax: 256-350-3873
info@lakeland.com
www.lakeland.com