

CERTIFICATE OF STERILITY

*Gamma Irradiation for CleanMax[®] Clean Sterile
Cleanroom Apparel*

Sterility Lot Number:	CS20220823-US
Product Release Date:	20220912
Minimum Specified Dose (kGy):	30.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	30.0
Maximum Delivered Dose (kGy):	45.7

Manufacturing Lot Number:	M2201077/1078
Sterilization Lot Number:	CR20220912

ITEM NUMBER(S)	SIZE(S)
CTL417CS	5X
CTL903CSP	LGXL

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.



M. LEE MORGAN, Ph.D., Global Director
Quality and Compliance