



Protect
Your
People®

CERTIFICATE OF STERILITY

Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

STERILITY LOT NUMBER
CS20221026-US

PRODUCT RELEASE DATE
20221107

STERILIZATION LOT	MIN Specified Dose	MAX Specified Dose	MIN Delivered Dose	MAX Delivered Dose
2201980	(kYg) 27.5	(kYg) 42.0	(kYg) 27.6	(kYg) 38.5


MANUFACTURING LOT NUMBER(S)	S2200316							
-----------------------------	----------	--	--	--	--	--	--	--

PRODUCT ITEM NUMBER(S)	SIZE(S)
CTL414CS	2X

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^{-6} . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.


M. LEE MORGAN, Ph.D.
Global Director, Quality and Compliance



Industrial Protective Clothing
Critical Environment
Fire Turnout Gear
Hi-Vis Safety Apparel

Lakeland Industries, Inc.

256-350-3873
800-645-9292

info@lakeland.com
www.lakeland.com