



Protect  
Your  
People®

## CERTIFICATE OF STERILITY

*Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel*

STERILITY LOT NUMBER  
**CS20221116-US**

PRODUCT RELEASE DATE  
**20230104**

STERILIZATION LOT	MIN Specified Dose	MAX Specified Dose	MIN Delivered Dose	MAX Delivered Dose
2300022-1	(kYg) 30.0	(kYg) 50.0	(kYg) 30.9	(kYg) 40.2


MANUFACTURING LOT NUMBER(S)	M2300008							
--------------------------------	----------	--	--	--	--	--	--	--

PRODUCT ITEM NUMBER(S)	SIZE(S)						
CTL903CSP	LG	XL					

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of  $10^{-6}$ . Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at [www.lakeland.com](http://www.lakeland.com) or contact Lakeland customer service at [info@lakeland.com](mailto:info@lakeland.com).

  
**M. LEE MORGAN, Ph.D.**  
Global Director, Quality and Compliance



Industrial Protective Clothing  
Critical Environment  
Fire Turnout Gear  
Hi-Vis Safety Apparel

Lakeland Industries, Inc.

256-350-3873  
800-645-9292

[info@lakeland.com](mailto:info@lakeland.com)  
[www.lakeland.com](http://www.lakeland.com)