



CERTIFICATE OF STERILITY

Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

		•					• • •	
STERILITY LOT NUMBER				PRODUCT RELEASE DATE				
CS20221202-US				20230104				
STERILIZATION LOT	MIN Specified Dose		MAX Specified Dose		MIN Delivered Dose		MAX Delivered Dose	
2300022-2	(kYg)	30.0	(kYg)	50.0	(kYg)	31.3	(kYg)	40.9
	•		•		-		-	
MANUFACTURING	M2300009							
LOT NUMBER(S)								
PRODUCT ITEM NUMBER(S)		SIZE(S)						
CTL903CSP		LG	XL					

Lakeland Industries, Inc., certifies that the Product/Item Number(s) and Lot Number(s) identified above have been processed and released in accordance with specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in **ANSI/AAMI/ISO** 11137-1: Sterilization of Heath Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10^-6. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at **info@lakeland.com**.

M. LEE MORGAN, Ph.D.

Global Director, Quality and Compliance



C S 2 O 2 2 1 2 O 2 - U S

Industrial Protective Clothing

Critical Environment Fire Turnout Gear **Lakeland Industries, Inc.**

256-350-3873 800-645-9292 info@lakeland.com www.lakeland.com

Hi-Vis Safety Apparel