



Lakeland Industries

202 Pride Lane SW
 Decatur, Alabama 35603
 Telephone: 256-350-3873
 Toll Free: 800-645-9291
 Fax: 256-350-0773

Certificate of Sterility

Gamma Irradiation for CleanMax Clean Sterile Cleanroom Apparel

Sterility Lot Number:	
Product Release Date:	
Minimum Specified Dose (kGy):	30.0
Maximum Specified Dose (kGy):	50.0
Minimum Delivered Dose (kGy):	
Maximum Delivered Dose (kGy):	

Manufacturing Lot Number:	
Sterilization Lot Number:	

Item numbers:	Size:

Lakeland certifies the product number and lot numbers above have been processed and released according to specifications and standard operating procedures. The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1: Sterilization of Health Care Products-Requirements for Validation and Routine Control-Radiation Sterilization to provide a Sterility Assurance Level of 10⁻⁶. Do not use if package integrity has been compromised.

Certificates of Sterility can be retrieved online at www.lakeland.com or contact Lakeland customer service at info@lakeland.com.

Approved by: Lakeland Industries
 Quality Assurance